

DECLARATION OF CONFORMITY

We declare under own responsibility that our

- **Approved Medical Scales**

7700, 7701, 7702, 7703, 7705, 7706, 7707, 7710, 7711,
7716, 7717, 7718, 7725, 7726, 6190, 7730, 7731, 7732

- **Indicators**

3020, 2790, 3010

manufactured by us meet the provisions of the
Council Directive **2009/23/EC** and MDD **93/42/EEC (2007/47/EC)**

In general we meet the following directives:

Directive / Standard	
Non automatic weighing instrument directive	2009/23/EC
Scale standard	EN45501
OIML ¹	R76
EMV directive	2004/108/EC
Standard for emission	EN60601-1-2
Standard for immunity	EN45501/60601-1-2
immunity criterion	A
ESD criterion	A
burst criterion	A
Surge criterion	A
Low voltage directive	2006/95/EC
Standard for electr. Safety	EN60601-1
Battery directive	91/157/EEC
Nickel directive	94/27/EEC
Medical device directive	93/42/EEC (2007/47/EC)
Food law (in Germany)	BgVO ⁵
Protection class	EN 60 529

¹) The decision to make an application for an OIML certificate should be made, in each case, when an application is made for a type approval


⁵) BgVO = Consumer Goods Order dated 07.01.98 (Federal Official Bulletin); "Components that can come into long-term contact with food." A confirmation from the supplier of the raw material that the material complies is sufficient

Please note:

This document annuls the previous sent Declaration of Conformity for non-approved medical scales and electronic height rods.

Validity: until 31.12.2011

Soehnle Professional GmbH & Co. KG
Manfred-von-Ardenne-Allee 12
D-71522 Backnang, Germany



Michael Schläger
General Manager

14.09.2010