

DECLARATION OF CONFORMITY

We declare under own responsibility that the following scales:

- *Non-approved Medical Scales*

6163, 7758, 7800, 7801, 7802, 7805, 7806, 7840, 7850,
8320, 8310, 9202, 9203, 2926, 9320, 9202, 9203, 7830

- *Electronic Height Rods*

5003, 5002.01, 5002.02, 5001.10, 5005.01

manufactured by us meet the provisions of the Council Directive

2004/108/EC

In general we meet the following directives:

Directive / Standard	
Company standard scale	FR 001 ²
EMV directive	2004/108/EC
Standard for emission	EN60601-1-2
Standard for immunity	EN60601-1-2
immunity criterion	A (3V/m) ⁴
ESD criterion	B
burst criterion	(B) ³
Surge criterion	(B) ³
Low voltage directive	2006/95/EC
Standard for electr. Safety	EN61010-1
Battery directive	91/157/EEC
Nickel directive	94/27/EEC
Food law (in Germany)	BgVO ⁵
Protection class	EN 60 529

²⁾ FR 001 = Factory Regulation 001

³⁾ in the event that electrical wires are present external to the product

⁴⁾ according to article 4 of the EMC act, a comment in German must be contained in the operating instructions, if the criterion cannot be maintained

⁵⁾ BgVO = Consumer Goods Order dated 07.01.98 (Federal Official Bulletin);
"Components that can come into long-term contact with food." A confirmation from the supplier of the raw material that the material complies is sufficient

Please note:

This document annuls the previous sent Declaration of Conformity for non-approved medical scales and electronic height rods.

Validity: until 31.12.2011

Soehnle Professional GmbH & Co. KG
Manfred-von-Ardenne-Allee 12
D-71522 Backnang, Germany

Michael Schläger
General Manager
14.09.2010